



EC - TYPE EXAMINATION CERTIFICATE

SK 08 - 001 MI-010 Rev. 1 Add. 0

This revision replaces all previous versions of this Certificate in full wording

Issued by **Slovenská legálna metrologia, n. o.** Notified body **1432**
Hviezdoslavova 31
974 01 Banská Bystrica
Slovak Republic

In accordance with Government Ordinance of the Slovak Republic No 294/2005 Coll. on measuring instruments as amended by Government Ordinance No 445/2010 Coll., which implements, in the Slovakia, the Directive 2004/22/EC of the European Parliament and of the Council of 31 March 2004 on measuring instruments as later amended (MID).

Applicable essential requirements Annex I and Annex MI-010 to MID

Manufacturer **ACTIA CZ s.r.o.**
Lesní 47
390 01 Tábor - Horky, Česká republika

Applicant Manufacturer

Measuring instrument **exhaust gas analyser**

Type **AT 505**
Trade mark **see Descriptive annex**
SW version 1.1.x or 1.2.x

Environment classes
- climatic (+ 5 to + 40) °C
- mechanical M2
- electromagnetic E2

Description and documentation The principal technical and metrological data, characteristics, instrument description and approval conditions are set out in the Descriptive annex to this EC - type examination certificate (11 pages), which is part of this EC - type examination certificate. The test reports, designs, schematic diagrams and documentation used during certification process are recorded under reference folder ACTIA_AT505_01_00.

Valid until **8 February 2018**

Date of issue 25 May 2015



Ing. Štefan Kral, PhD.
Representative of Notified Body



Where the instrument is subject to other Directives covering other aspects, this EC - type examination certificate is valid, assuming that the instrument conforms to the provisions of those Directives. Without written permission of the notified body this certificate may be reproduced only as a whole.